


P. 1061
St



TUCSON POLICE DEPARTMENT
CRIME LABORATORY

INTOXILYZER SERVICE RECORD

Date: 3/24/04
Intoxilyzer Serial Number: 66-004015 (#18)
QAS: Irene Doran (46331)

Problem Description:

DVM below DPS range (.200 - .600) on

3/14/04.

If the instrument is taken out of service, complete the following:

Date out of service 3/24/04

Calibration Check completed prior to removal from service:

☒ Yes Date: 3/14/04 ☐ No

SQAP completed prior to removal from service:

☒ Yes Date: 3/14/04 ☐ No

Date returned to service: 5/24/04

Corrective Action Taken or Repairs: Pulled from service on 3/24/04
Sent to Applied Electronics for DVM repair. Received
from Applied Electronics with printer problems. Instrument
returned to Applied Electronics for further repair

Tuesday May 25, 2004

To Whom It May Concern:

Intoxilyzer **66-004015** was taken out of service for repair on **3/24/04**.

I cannot testify that it was working properly and accurately between the dates of
3/14/04 and **3/24/04**.

This instrument was located at the **ODM** substation **3201 E First Ave.**

Irene Doran (46331)
Intoxilyzer 5000 QAS
Tucson Police Department Crime Lab
791-4494